


GUIDE SEITRON

Accendere lo strumento ed attendere la fine dell'autozero.


Collegare lo strumento al PC tramite il cavo USB in dotazione.




The screenshot shows the Seitron website's 'Novo' section. At the top, there is a navigation bar with the Seitron logo and links: PRODOTTI, SHOP, AZIENDA, ASSISTENZA, DOWNLOAD, CONTATTI, and MAGAZINE. Below the navigation bar is a large image of industrial equipment with a dropdown menu open, showing 'App e Software', 'Catalogo, Listini e Brochures', and 'Manuali e Firmware analizzatori' (highlighted with a red arrow). To the right of the image, text instructs users to go to www.seitron.it and select 'DOWNLOAD > Manuali e Firmware analizzatori'. Below the image, the 'Novo' section is displayed, featuring a sidebar with a smartphone icon and a main content area with four links: 'Manuale' (with a 'Novo' link below it), 'Guida rapida' (with a 'Novo' link below it), 'Firmware' (with a 'Novo FW Updater 2.3.5' link below it, highlighted with a red arrow), and 'Smart Analysis Software' (with a 'Versione 4.2.0.1' link below it). At the bottom of the main content area, there is a section for 'Smart Analysis App' with links for 'Versione Android', 'Versione iPhone', and 'Versione iPad'. A small footnote at the bottom left states: '* Compatibile con i sistemi Windows 7 e successivi aggiornamenti'.

Accedere al sito www.seitron.it e selezionare
DOWNLOAD > Manuali e Firmware analizzatori

Avviare il download del software **NOVO FW UPDATER**




Accedere alla cartella zip *NovoUpdater* appena scaricata e avviare l'installazione del software (doppio clic sull'applicazione *NovoUpdater.exe*)



Se appare il seguente messaggio, proseguire selezionando:

Ulteriori informazioni

Esegui comunque



Avviare l'aggiornamento e attendere fino a completamento

